Continuous Monitoring Final Report

METEC
METHANE EMISSIONS TECHNOLOGY EVALUATION CENTER

Performer: SENSIT
Report Generated: July 11, 2024

1 Experiment Summary

1.1 Experiment Demographics

Testing included a total of 351 experiments performed between 2024/02/06 00:00 and 2024/04/29 00:00. A total of 783 controlled releases were performed during these experiments. Individual experiments included between 1 and 5 controlled releases, with an average of 2.23 controlled releases per experiment (Figure 1). Figure 2 shows the emission rate of controlled releases during experiments. Figure 3 shows a histogram of the duration of controlled releases performed. Emission rates ranged from 2.3 to 189.0 slpm whole gas. The variation of controlled release rates (kg CH_4/h) with experiment duration (h) is captured in Figure 4. Each controlled release is classified as either a True positive (TP) or False negative (FN). Figure 5 summarizes the location of controlled releases by equipment unit. The range of controlled release rates (kg CH_4/h) and detection classifications across the equipment groups involved in the experiments is illustrated in Figure 6.



Figure 1: Histogram of number of controlled releases included in each experiment



Figure 2: Histogram of metered emission rate (kg CH_4/h) for all controlled releases performed during experiments. Each controlled release is classified as true positive (TP) or false negative (FN).



Figure 3: Histogram of controlled release duration (h) for all controlled releases performed during experiments. Each controlled release is classified as true positive (TP) or false negative (FN).



Figure 4: Controlled release rate (kg CH_4/h) versus controlled release duration (h). Markers indicate True Positive (TP) and False Negative (FN) detection classification. Data are summarized by histogram shown parallel to the respective axes.



Figure 5: Distribution of controlled releases across equipment units. Each controlled release is classified as true positive (TP) or false negative (FN).



Figure 6: Controlled release rate (kg CH_4/h) versus equipment group. The whiskers of each boxplot represent the range of all controlled releases metered in each equipment group. While the boxes represents 50% of data points about the median controlled release rate

1.2 Meteorology

Meteorological data during testing is summarized in the following figures. Figure 7 shows the average ambient temperature during each experiment. Temperature ranged from -11.0 ^{o}C to 23.4 ^{o}C, with a mean of 5.71 ^{o}C. Figure 8 shows the average wind speed of each experiment. Average wind speed ranged from 0.516 m/s to 12.9 m/s, with a mean of 2.9 m/s. Figure 9 shows the average direction during each experiment. Figure 10 shows the average wind speed and direction of each experiment as a wind rose.



Figure 7: Histogram of average temperature (^{o}C) during each experiment



Figure 8: Histogram of average wind speed (m/s) during each experiment



Figure 9: Histogram of average wind direction during each experiment



Figure 10: Experiment wind rose. Wedges show percent of experiments by average wind speed and direction. 0^o represents the true North of the test center.

2 Performance Metrics

Metrics as described in section 6 in the protocol are reported in this section. Primary metrics are reported for all performers. Secondary metrics are reported only if the performer detection reports included the required data for their calculation.

2.1 Primary Metrics

2.1.1 Classification of Detections

SENSIT reported a total of 666 detection reports including 622 unique EmissionSourceIDs. 13 reported sources were excluded from classification as they may indicate an emission from the test center during a maintenance period or another test conducted at the test center that was not following the continuous monitoring protocol. Remaining detection reports were matched to controlled releases following the method outlined in the test protocol to identify true positives, false positives, and false negatives. The classification of controlled releases and detections is shown in Table 1. The classification of individual detections and controlled releases can be found in the accompanying data files.

Table 1: Classification of controlled releases and detection reports.

Level	True Positive	False Negative	False Positive	Excluded from Analysis	Total
Controlled Releases	560	215	-	8	783
Detection Reports	560	-	49	13	622

2.1.2 Probability of Detection

The probability of detection (POD) curves derived from the classified detection and controlled release records is illustrated in Figure 11. The upper subplot of the figure shows a logistic regression performed against the true positive and false negative results. The lower subplot shows the POD calculated using a variable bin width each containing approximately equal number of data points. The bins are obtained through quantile-based discretization limiting number of points per bin to between 30 to 50. Power function with intercept set to zero is used to fit calculated POD data points. In this approach, the statistical significance of each bin, is approximately equal. Figure 12 shows the probability of detection curves for controlled release rates normalized by wind speed.



Figure 11: Probability of detection versus emission rate. Upper subplot is the probability of detection vs emission rate (kg CH_4/h) assessed with logistic regression including all true positives (equipment unit-, equipment group-, and facility-levels). True positive and false negative detections are shown with markers at y = 1 and y = 0 respectively. The regression is performed on bootstrapped results to show a cloud of curves to illustrate uncertainty in the result. The lower subplot shows the probability of detection versus emission rate (kg CH_4/h) fitted using power functions. The x-axis is divided into equal-sized bins with each marker (pod) as the fraction of controlled releases in a bin classified as TPs. The emission rate at which the POD reaches 90% is shown. Each pod data point is bootstrapped to produce a cloud of curves illustrating associated uncertainty. When the bootstrapping could not evaluate the lower and upper empirical confidence limit on the emission rate of 90% POD, they are given as 0 and NA respectively. Curve fits (dotted colored lines) obtained using other (5-16) quantile-based discretizations are shown for comparison. The mean count of points per bin along with the minimum and maximum counts across all bins is also shown in the figure.



Figure 12: Probability of detection versus wind normalized emission rate. Upper subplot is the probability of detection vs wind normalized emission rate $(kg\ CH_4/h)/(m/s)$ assessed with logistic regression including all true positives (equipment unit-, equipment group-, and facility-levels). True positive and false negative detections are shown with markers at y = 1 and y = 0 respectively. The lower subplot shows the probability of detection versus wind normalized emission rate (kg CH_4/h) fitted using power functions. The x-axis is divided into equal-sized bins with each marker (pod) as the fraction of controlled releases in a bin classified as TPs. The wind normalized emission rate at which the POD reaches 90% is shown. Each pod data point is bootstrapped to produce a cloud of curves illustrating associated uncertainty. When the bootstrapping could not evaluate the lower and upper empirical confidence limit on the wind normalized emission rate of 90% POD, they are given as 0 and NA respectively. Curve fits (dotted colored lines) obtained using other (5-16) quantile-based discretizations are shown for comparison. The mean count of points per bin along with the minimum and maximum counts across all bins is also shown in the figure.



Figure 13: The left subplot shows the probability of detection vs emission rate (kg CH_4/h) calculated by binning data along the x-axis. Separate curves are illustrated for equipment unit, equipment group, and facility levels. Markers represent mean emission rate and observed probability of detection within each bin. X whiskers indicate maximum and minimum emission rate in each bin. Y whiskers indicate maximum and minimum probability of detection when empirical data is bootstrapped. The number of data points within each bin is plotted on right hand axis. The right subplot shows the probability of detection vs wind normalized emission rate ($kg\ CH_4/h)/(m/s$) calculated by binning data along the x-axis. Separate curves are illustrated for equipment unit, equipment group, and facility levels. Markers represent mean emission rate and observed probability of detection within each bin. X whiskers indicate maximum and minimum wind normalized emission rate in each bin. Y whiskers indicate maximum and minimum probability of detection when empirical data is bootstrapped. The number of data points within each bin is plotted on right hand axis.

2.1.3 False Positive Fraction

The false positive fraction ($N_{FP}/(N_{FP} + N_{TP})$) derived from all detections was 0.0805.



Figure 14: Classification of performer detection alerts. True Positive (TP) alerts were paired with a controlled release at the test center. False Positive (FP) alerts were unpaired.



Figure 15: Categorization of false positives. The pie chart shows the fraction of FPs due to no controlled release and excess detection alerts sent during experiments.

2.1.4 False Negative Fraction

The false negative fraction ($N_{FN}/(N_{FN} + N_{TP})$) derived from all controlled releases was 0.277.



Figure 16: Classification of controlled releases. True Positive (TP) releases were paired with a detection reported by the performer. False Negative (FN) releases were unpaired.

2.1.5 Detection Time

The detection time of individual detections is illustrated in figure 17. The detection time ranged from 22.664 to 207.796 hours. The 2.5% and 97.5% percentiles of the detection time were 44.717 and 174.178 hours respectively. The mean detection time was 128.642 hours. Note, the detection time may only be calculated for true positive detections.



Figure 17: Detection time was calculated as the time difference between the start time of a controlled release and the time when the matched detection was received by the test center.

2.2 Secondary Metrics

2.2.1 Localization Precision (Equipment Unit)

Table 2 lists the number of true positives at the equipment unit, equipment group, and facility-level.

Table 2: Localization Precision (Equipment Unit)

Level	True Positive Count
Equipment Unit	143
Equipment Group	364
Facility	53



Figure 18: Categorizing True Positive detections by localization precision levels. The localization precision levels include the Equipment unit level, Equipment group level, and the Facility level.



Figure 19: Categorizing the classification of controlled releases and equipment localization precision based on the range of release rate. Controlled releases are classified as True Positives (TP) and False Negatives (FN). The localization precision levels include the Equipment unit level, Equipment group level, and the Facility level.

2.2.2 Localization Accuracy (Equipment Unit)

Table 3 lists the localization accuracy (the fraction of reports identified as true positive) at each level of precision. Precision levels are the equipment unit, the equipment group or better (i.e. equipment unit + equipment group), and facility or better (i.e. equipment unit + equipment group + facility). The fraction of reports identified as false positive is also included in the table.

Table 3: Localization Accuracy (Equipment Unit)

Level	Localization Accuracy
Equipment Unit	0.235
Equipment Group	0.833
Facility	0.92
False Positive Fraction	0.0805

2.2.3 Operational Factor

The operational factor (1 - $\sum t_{offline}$ / t_{total}) derived from all offline reports received during testing was 1.0.

2.2.4 Quantification Accuracy (Absolute)

SENSIT did not report data required to compute this metric.

2.2.5 Quantification Accuracy (Relative)

SENSIT did not report data required to compute this metric.

2.2.6 Quantification Precision (Absolute)

SENSIT did not report data required to compute this metric.

2.2.7 Quantification Precision (Relative)

SENSIT did not report data required to compute this metric.

2.2.8 Localization Accuracy (Single Coordinate)

SENSIT did not report data required to compute this metric.

2.2.9 Localization Accuracy (Bounding Box)

SENSIT did not report data required to compute this metric.

2.2.10 Bounding Box Accuracy

SENSIT did not report data required to compute this metric.

2.2.11 Localization Precision (Bounding Box)

SENSIT did not report data required to compute this metric.

3 Documentation of Test Protocol

A copy of the test protocol is provided in Continuous Monitoring Protocol R1.0.pdf in the zip folder with this report.

4 Documentation of System Under Test

Table **??** lists the documentation of the system under test submitted by SENSIT. SENSIT deployed 14 sensors at a sensor density of 604.0 m^2 per sensor. Data left empty was reported as N/A.

Table 4: Performer Information

Company	SENSIT
Product Name	SPOD (MOS Version)
(1) Location (latitude, longitude, height) of system components including central instruments, meteorological station, retroreflectors, sensor nodes, or any other equipment installed at or near the Test Center.	Height on each unit is approximately 5 feet. There is a 3D MET on 1603, Mechanical METs on 1027, 1042, and 1001. There are ulatrasonic METs on 1611 and 1610. The following link will direct you to a map with precise coordinates (Lat/Long): https://maps.app.goo.gl/ikmnX9ptLcMs9bwPA
(2) Model number of each component in (1)	1027, 1605, 1607, 1603, 1042, 1610, 1602, 1606, 1041 (aka 1017), 1608, 1609, 1001, 1611, and 1604.
(3) Power configuration of each component in (1)	Power Requirement Built in Rechargeable Battery Maintained by Solar Panel Power Consumption Less than 2 Watts
(4) Revision number of software installed in each component in (1) that includes performer-specific software components, revisions, or customizations.	firmware for 1027, 1042, 1041 (aka 1017), and 1001 is r-126. Firmware for 1605, 1607, 1603, 1610, 1602, 1606, 1608, 1609, 1611, and 1604 is mos-hdr.
(5) Revision number of any software analytics installed offsite.	The software platform that we are using does not have a revision. It is currently under BETA. It is an JavaScript API pulled from sensitconnect.net
(6) Confidence level at which emission detection data are reported.	We are confident in the site that is leaking but not the specific asset. Everything we are reporting is just a start and end time with a general site identifier (eg. We list 4T-2 for every tank leakwe do not try to differentiate between 4T-1, 4T-2, and 4T-3). This goes for all assets. We are leaving all of the other paramaters blank in the report. There is no quantification.

5 Controlled Release and Detection Data

Controlled release and classification data are provided in classifiedReports.csv in the zip folder with this report. The raw data reported by the performer is provided in detectionReports.xlsx in the zip folder with this report.

6 Offline Reports

Offline reports are provided in offlineReports.csv in the zip folder with this report.

7 Flow Meter Calibrations

Meter calibrations are available from the test center by request.